

AD-A111 390 CALIFORNIA UNIV LOS ANGELES INST OF GEOPHYSICS AND --ETC F/G 8/14
DESIGN AND FABRICATION REPORT.(U)
SEP 81 R C SNARE


N00014-77-C-0343
NL

UNCLASSIFIED

1 - 1
OF 1



END
DATE
FILMED
3 82
DTIC



NR 323-044/3/11/77(427)

(12)

ADA111390

Design and Fabrication Report

by

Robert C. Snare

Institute of Geophysics and Planetary Physics

University of California

Los Angeles, California 90024

Final Contract Report

1 March 1977 - 30 September 1981

Approval for public release,
distribution unlimited.

DTIC
ELECTE
FEB 26 1982
A

Office of Naval Research

Department of the Navy

800 N. Quincey Street

Arlington, Virginia 22217

DTIC FILE COPY

Unclassified

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

REPORT DOCUMENTATION PAGE		READ INSTRUCTIONS BEFORE COMPLETING FORM
1. REPORT NUMBER NR323-044/3/11/77(427)	2. GOVT ACCESSION NO. AD-A114390	3. RECIPIENT'S CATALOG NUMBER
4. TITLE (and Subtitle) Design and Fabrication Report		5. TYPE OF REPORT & PERIOD COVERED Final Report 1 March 1977 - 30 Sept. 1981
7. AUTHOR(s) Robert C. Snare		6. PERFORMING ORG. REPORT NUMBER None
9. PERFORMING ORGANIZATION NAME AND ADDRESS Institute of Geophysics & Planetary Physics University of California Los Angeles, CA 90024		8. CONTRACT OR GRANT NUMBER(s) ONR N00014-77-C-0343
11. CONTROLLING OFFICE NAME AND ADDRESS Office of Naval Research, Code 427 800 N. Quincey Street Arlington, Virginia 22217		10. PROGRAM ELEMENT, PROJECT, TASK AREA & WORK UNIT NUMBERS NR323-004/3/11/77 (427)
14. MONITORING AGENCY NAME & ADDRESS (if different from Controlling Office)		12. REPORT DATE
		13. NUMBER OF PAGES 5
		15. SECURITY CLASS. (of this report) Unclassified
		15a. DECLASSIFICATION/DOWNGRADING SCHEDULE
16. DISTRIBUTION STATEMENT (of this Report) A-Approval for public release; distribution unlimited.		
17. DISTRIBUTION STATEMENT (of the abstract entered in Block 20, if different from Report)		
18. SUPPLEMENTARY NOTES		
19. KEY WORDS (Continue on reverse side if necessary and identify by block number) Magnetometer, data system.		
20. ABSTRACT (Continue on reverse side if necessary and identify by block number) The design and fabrication of fluxgate magnetometers and data systems is described.		

DD FORM 1 JAN 73 1473

EDITION OF 1 NOV 65 IS OBSOLETE
S/N 0102-LF-014-6601

Unclassified

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

Table of Contents

1. Contract objectives
2. Fluxgate magnetometer
3. Digital data system
4. Construction



Preparation For	
DTIC DTARI	<input checked="" type="checkbox"/>
DTIC TAB	<input type="checkbox"/>
Unannounced	<input type="checkbox"/>
Justification	
Distribution/	
Availability Codes	
Avail and/or	
Spec	Special
A	

1. Contract Objectives

The purpose of the contract was to develop, fabricate and test seven triaxial fluxgate magnetometer and digital data systems, for use in studying geomagnetic phenomenology.

2. Fluxgate Magnetometer

The triaxial fluxgate magnetometer followed the design of Power [1] with modifications. The modifications consisted of changing all digital functions in the field offset system from hardwired logic to microprocessor control.

The fluxgate sensors were supplied by Naval Surface Weapons Center, White Oak, Silver Spring, MD. The design was similar to that reported by Gordon et al. [2]. An additional winding was added to the sensors to provide offset field capability and extend the range of the sensors to $\pm 65,000$ nT.

3. Digital Data System

The digital data system is controlled by a RCA 1802 microprocessor. The microprocessor was programmed to control the analog to digital convertor functions, to average data and to format the serial output data stream.

The system digitized the three magnetometer signals and three auxillary inputs at 240 times per second. The data from each channel was then averaged to obtain 60 samples per second for each channel. These six main data channels are formatted together with sixteen low rate channels.

Each sample is converted to a sixteen bit digital word. The serial output data passes through a universal asynchronous receiver transmitter circuit (UART) and a modem circuit. The output data rate is 9600 bits per second. The data is carried by a coaxial cable that also carries direct current power and command information to the system. A data interface unit was developed that received data from the magnetometer units and interfaced the data to a computer for processing and storage.

3. Construction

All of the electronic circuits were packaged on printed circuit cards measuring 11.4 x 15.2 cm (4.5 x 6 in). The circuit cards that formed the microprocessor plugged into a standard digital data buss that carried address and data lines for the unit. The cards were housed in a chassis measuring 13.3 x 22.8 x 77.1 cm (5.25 x 9 x 30.35 in). A total of seven units were fabricated, tested and delivered during the contract.

References

- [1] J.J. Power, A digital offset fluxgate magnetometer for use in remote geomagnetic observatories, AFCRL-TR-0603, 1973.
- [2] D.I. Gordon, R.H. Lundston, R.A. Chiarodo and H.H. Helms, Jr., A fluxgate sensor of high stability for low field magnetometry, IEEE Transactions on Magnetism, Vol. MAG-4, No. 3, September 1968.

DATE
FILMED
- 8